



Contribution ID: 48

Type: Contributed e-poster

Toward sub-100 ps ToF-PET systems employing the FastIC ASIC with analog SiPMs

Tuesday, July 4, 2023 4:35 PM (1h 25m)

Pushing the limits of the readout system in standard ToF-PET scanners will improve tomographic image reconstruction quality. Scintillation crystals optically coupled to photo-detectors, together with fast readout electronics, are the key elements to improve Coincidence Time Resolution (CTR). The 8-channel FastIC ASIC is capable of processing fast signals with a precise time stamp and a linear energy measurement with ≈ 12 mW power consumption per channel.

This work provides a complete analysis of the performance of the FastIC ASIC coupled to different scintillators and photo-detectors from different manufacturers, giving special attention to FBK sensors.

Primary author: PORTERO, Sara (ICCUB)

Co-authors: MARISCAL, Antonio (Institut de Ciències del Cosmos UB); SILVA, Juan José (ICCUB); MANERA, Rafel (ICCUB)